proceedings

MDPI

Abstract

A Phenylalanine Ammonia Lyase Capacitive Sensor for Phenylalanine Detection †

Bruno Andò [1,*], Salvatore Castorina [1], Ludovica Maugeri [1], Salvatore Petralia [2], Maria Anna Messina [3], Martino Ruggieri [4], Giovanni Neri [5], Angelo Ferlazzo [5], Emilio Sardini [6] and Mauro Serpelloni [6]

1 DIEEI, University of Catania, 95123 Catania, Italy; salvatore.castorina@unict.it (S.C.); ludovica.maugeri@unict.it (L.M.)
2 Department of Drug and Health Sciences, University of Catania, 95123 Catania, Italy; salvatore.petralia@unict.it
3 Expanded Newborn Screening Laboratory, Unit of Clinical Pediatrics, University of Catania, 95123 Catania, Italy; mmessina@unict.it
4 Unit of Clinical Pediatrics, Department of Clinical and Experimental Medicine, University of Catania, 95124 Catania, Italy; m.ruggieri@unict.it
5 Department of Engineering, University of Messina, 98122 Messina, Italy; giovanni.neri@unime.it (G.N.); angelo.ferlazzo@unime.it (A.F.)
6 Department of Information Engineering, University of Brescia, 25123 Brescia, Italy; emilio.sardini@unibs.it (E.S.); mauro.serpelloni@unibs.it (M.S.)
* Correspondence: bruno.ando@unict.it
† Presented at the XXXV EUROSENSORS Conference, Lecce, Italy, 10–13 September 2023.

Abstract: In this paper, an easy-to-use and fast biosensor for phenylalanine quantification in patients affected by phenylketonuria is investigated. The phenylalanine concentration was indirectly estimated through the ammonia released as a by-product of an enzymatic reaction, which was then detected by exploiting an yttria-stabilized zirconia layer deposited over an interdigitated capacitive sensor. The latter was manufactured by rapid-prototyping technologies. A sensor limit of detection higher than 1.25 μM was estimated, along with an accuracy better than 18.31 μM.

Keywords: phenylalanine; ammonia; biosensor; interdigitated capacitor; rapid prototyping

check for updates

Citation: Andò, B.; Castorina, S.; Maugeri, L.; Petralia, S.; Messina, M.A.; Ruggieri, M.; Neri, G.; Ferlazzo, A.; Sardini, E.; Serpelloni, M. A Phenylalanine Ammonia Lyase Capacitive Sensor for Phenylalanine Detection. *Proceedings* **2024**, *97*, 51. https://doi.org/10.3390/proceedings2024097051

Academic Editors: Pietro Siciliano and Luca Francioso

Published: 19 March 2024

Copyright: © 2024 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (https://creativecommons.org/licenses/by/4.0/).

1. Introduction

Phenylketonuria (PKU) is a genetic disorder of phenylalanine (Phe) metabolism that may potentially lead to severe neurological damages. To date, a specific diet and monitoring the Phe levels in the blood is the unique therapy available [1]. Commonly, techniques for Phe quantification are complex, expensive and time-consuming [2]. This research work addresses the development of a low-cost and fast-responsive capacitive sensor for Phe quantification, which offers an accuracy compliant with early warning tasks. The sensing methodology exploits the indirect estimation of Phe by detecting aqueous ammonia (NH_3(aq)) as a by-product of a specific enzymatic reaction with Phenylalanine Ammonia Lyase (PAL). Most NH_3 sensors are aimed at gas detection [3], while few approaches address the quantification of NH_3(aq) [4], none of which provides solutions for measurements in small volumes. The modeling and design flow of the sensor are discussed in [5], while the ability of YSZ to detect NH_3(aq) is demonstrated in [6]. In this paper, the behavior of the complete sensor is investigated, including the Phe-PAL enzymatic reaction. The main outcomes of this approach include: (i) a capacitive readout strategy, which is convenient for low analyte concentrations; (ii) high specificity to the target analyte, thanks to the enzymatic reaction; (iii) faster response times compared to those of traditional Phe quantification methods; (iv) low costs, also provided by the adopted rapid-prototyping technologies.

2. The Developed Device

The structure and real view of the capacitive sensor are shown in Figure 1a,b. The IDC electrodes were realized by aerosol jet printing a gold-based conductive ink over a 125 μm thick Polyether Ether Ketone (PEEK) substrate. The sensor active area was 13.5 × 9.8 mm^2, and the IDC electrodes were 500 μm wide, with a spacing of 300 μm. A 50 μm thick yttria-stabilized zirconia (YSZ) dielectric functional layer, sensitive to NH_3(aq), was spray-coated over the IDC electrodes. The YSZ layer was subjected at room temperature, for 45 min, to an O_3/UV_{254nm} process performed by BioForce equipment (Nanosciences) to increase its surface hydrophilicity (126.6° ± 3°). The proposed enzymatic method exploits the deamination of Phe catalyzed by the PAL enzyme to produce trans-cinnamic acid and NH_3. The yield of the latter is proportional to the Phe concentration [7]. PAL showed optimal sensitivity at temperatures ranging from 32 to 40 °C and pH values ranging from 7 to 8. A rigid supporting frame for the IDC sensor was milled in 1.55 mm thick FR4. The frame also provided the reaction chamber hosting the MUT.



Figure 1. The sensor: (**a**) top view (layout), (**b**) cross section and real view, (**c**) calibration diagram.

3. Experimental Results

The sensor behavior was investigated by observing its response to standard test solutions with Phe concentrations in the range of [0, 800] μM. The test solutions were prepared by mixing Phe with 10 μL of PAL (containing 0.156 units in 200 μL of 10 mM sodium phosphate buffer at pH 8.31) and 1 mL of 2 mM sodium phosphate buffer at pH 7.5. A volume of 500 μL of these solutions was dropped on the YSZ sensing layer, and the temperature was set to 37 °C (±2 °C) for 30 min to realize the enzymatic reaction. In each trial, the sensing chamber was completely filled. Time evolution of the sensor capacitance was acquired by a GW Instek LCR-6300 precision LCR meter. The obtained calibration diagram is shown in Figure 1c. A piecewise linear model was used to interpolate the data. The sensor responsivity was 5.31 × 10^{-1} pF/μM for Phe concentrations in the [0–200] μM range and 2.86 × 10^{-1} pF/μM for Phe concentrations in the [200–800] μM range. The accuracy in the 3σ limit, for the two considered intervals, resulted to be 18.31 μM and 12.34 μM, while the estimated limit of detection was 0.98 μM and 1.25 μM, respectively.

Author Contributions: Conceptualization and methodology: B.A. and S.P.; validation: S.C., L.M., M.A.M. and M.R.; development: S.C., A.F., G.N., E.S. and M.S. All authors have read and agreed to the published version of the manuscript.

Funding: This research was funded by 1.1.5 PO-FESR 2014–2020, project "PKU", G89J18000710007.

Institutional Review Board Statement: Not applicable.

Informed Consent Statement: Not applicable.

Data Availability Statement: No new data were created.

Acknowledgments: This work was supported by the project "PKU", G89J18000710007, 1.1.5 PO-FESR 2014–2020.

Conflicts of Interest: The authors declare no conflict of interest.

References

1. Blau, N.; Hennermann, J.B.; Langenbeck, U.; Lichter-Konecki, U. Diagnosis, classification, and genetics of phenylketonuria and tetrahydrobiopterin (BH4) deficiencies. *Mol. Genet. Metab.* **2011**, *104*, S2–S9. [CrossRef] [PubMed]
2. da Silva, K.P.; Ptak, M.; Pizani, P.S.; Mendes Filho, J.; Melo, F.E.A.; Freire, P.T.C. Raman spectroscopy of L-phenylalanine nitric acid submitted to high pressure. *Vib. Spectrosc.* **2016**, *85*, 97–103. [CrossRef]
3. Shen, W.C.; Shih, P.J.; Tsai, Y.C.; Hsu, C.C.; Dai, C.L. Low concentration ammonia gas sensors manufactured using the CMOSMEMS technique. *Micromachines* **2020**, *11*, 92. [CrossRef] [PubMed]
4. Winquist, F.; Spetz, A.; Lundström, I.; Danielsson, B. Determination of ammonia in air and aqueous samples with a gas-sensitive semiconductor capacitor. *Anal. Chim. Acta* **1984**, *164*, 127–138. [CrossRef]
5. Andò, B.; Baglio, S.; Castorina, S.; Graziani, S.; Messina, M.; Petralia, S.; Tondepu, S.V.G. A Capacitive Readout Strategy for Ammonia Detection: Design Flow, Modeling and Simulation. In Proceedings of the 2021 IEEE Sensors Applications Symposium (SAS), Sundsvall, Sweden, 23–25 August 2021; pp. 1–6.
6. Andò, B.; Baglio, S.; Castorina, S.; Graziani, S.; Tondepu, S.V.G.; Petralia, S.; Messina, M.A.; Maugeri, L.; Neri, G.; Ferlazzo, A. A Capacitive Sensor, Exploiting a YSZ Functional Layer, for Ammonia Detection. *IEEE Trans. Instrum. Meas.* **2022**, *71*, 9505811. [CrossRef]
7. Messina, M.A.; Meli, C.; Conoci, S.; Petralia, S. A facile method for urinary phenylalanine measurement on paper-based lab-on-chip for PKU therapy monitoring. *Analyst* **2017**, *142*, 4629–4632. [CrossRef] [PubMed]

Disclaimer/Publisher's Note: The statements, opinions and data contained in all publications are solely those of the individual author(s) and contributor(s) and not of MDPI and/or the editor(s). MDPI and/or the editor(s) disclaim responsibility for any injury to people or property resulting from any ideas, methods, instructions or products referred to in the content.