

Available online at www.sciencedirect.com

## **ScienceDirect**

journal homepage: www.elsevier.com/locate/gie



## ORIGINAL ARTICLE/ARTICOLO ORIGINALE

CONGRESSO INTERNAZIONALE PARMA 2014 VINCITORE PREMIO RICCARDO GARBEROGLIO

# Effect of cyclic torsional preloading on cyclic ( fatigue resistance of ProTaper Next and Mtwo nickel—titanium instruments



Influenza della torsione sulla resistenza alla fatica ciclica di strumenti in Ni-Ti

Eugenio Pedullà<sup>a,\*</sup>, Fabio Lo Savio<sup>b</sup>, Gianluca Plotino<sup>c</sup>, Nicola Maria Grande<sup>c</sup>, Silvia Rapisarda<sup>a</sup>, Gianluca Gambarini<sup>c</sup>, Guido La Rosa<sup>b</sup>

<sup>a</sup> Department of Surgery, University of Catania, Catania, Italy

<sup>b</sup> Department of Industrial Engineering, University of Catania, Catania, Italy

<sup>c</sup> Department of Endodontics, 'Sapienza' University of Rome, Rome, Italy

Received 7 March 2015; accepted 23 April 2015 Available online 13 June 2015

### **KEYWORDS**

Cyclic fatigue resistance; Torsional preload; NiTi fracture; M-wire; Torsional fracture.

#### Abstract

Aim: To evaluate the influence of different torsional preloads on cyclic fatigue resistance of endodontic rotary instruments made by conventional nickel—titanium (NiTi) or M-Wire. *Methods:* Eighty new Mtwo (#25/0.06) (Sweden & Martina, Due Carrare, Padova, Italy) and ProTaper Next X2 (Dentsply Maillefer, Ballaigues, Switzerland) were used. The Torque and distortion angles at failure of new instruments (n = 10) were measured and 0% (n = 10), 25%, 50%, and 75% (n = 20) of the mean ultimate torsional strength as preloading condition were applied according to ISO 3630-1 for each brand. The twenty files tested for every extent of preload were subjected to 25 or 50 torsional cycles (n = 10). After torsional preloading, the number of cycles to failure was evaluated in a simulated canal with 60° angle of curvature and 5 mm of radius of curvature. Data were analyzed using 2-way analysis of variance. Scanning electron microscope (SEM) was performed to evaluate the fracture surface of each fragment.

\* Corresponding author at: Via Cervignano 29, 95129 Catania, Sicily, Italy. Tel.: +39 339 2613264. *E-mail*: eugeniopedulla@gmail.com (E. Pedullà).

Peer review under responsibility of Società Italiana di Endodonzia.



http://dx.doi.org/10.1016/j.gien.2015.05.002

1121-4171/© 2015 Società Italiana di Endodonzia. Production and hosting by Elsevier B.V. All rights reserved.

*Results:* ProTaper Next X2 reduced their cyclic fatigue after 25 or 50 cycles of every torsional preloading (25%, 50%, and 75%) (P < 0.01). Cyclic fatigue of Mtwo was reduced by 50 cycles of every torsional preloading and only after the 75% of preload for 25 cycles (P < 0.01).

*Conclusions:* Torsional preloads reduced the cyclic fatigue resistance of M-wire and conventional (as ProTaper Next and Mtwo) NiTi rotary instruments except for Mtwo with 25% or 50% of torsional preloading.

© 2015 Società Italiana di Endodonzia. Production and hosting by Elsevier B.V. All rights reserved.

#### PAROLE CHIAVE

Resistenza alla fatica ciclica; precarico torsionale; frattura strumenti in Nichel-Titanio; M-wire; frattura torsionale.

#### Riassunto

*Obiettivo:* Scopo del lavoro è stato valutare l'influenza di diversi precarichi torsionali sulla resistenza a fatica ciclica degli strumenti rotanti endodontici in nickel-titanio (NiTi) convenzionale o M-Wire.

*Metodi:* Sono stati utilizzati ottanta Mtwo (# 25/0,06) (Sweden & Martina, Due Carrare, Padova, Italia) e ProTaper Next X2 (Dentsply Maillefer, Ballaigues, Svizzera): La coppia e l'angolo di distorsione necessari per la frattura di strumenti nuovi (n = 10) sono stati misurati e un valore pari a 0% (n = 10), 25%, 50% e 75% (n = 20) della resistenza torsionale media finale è stata applicata come condizione di precarico secondo la norma ISO 3630-1 per ogni marca. I venti file testati per ogni tipologia di precarico sono stati sottoposti a 25 o 50 cicli torsionali (n = 10). Dopo precarico torsionale è stato valutato in un canale simulato, con 60° di angolo di curvatura e 5 mm di raggio di curvatura, il numero di cicli a frattura (NCF). I dati sono stati analizzati mediante analisi della varianza a 2 vie. Microscopia elettronica a scansione (SEM) è stata eseguita per valutare la superficie di frattura di ciascun frammento.

*Risultati:* I ProTaper Next X2 hanno ridotto la loro resistenza a fatica ciclica dopo 25 o 50 cicli di ogni precarico torsionale (25%, 50%, e il 75%) (P < .01). La resistenza a fatcica cilcica degli strumenti Mtwo è stata ridotta da 50 cicli di ogni precarico torsionale e solo dopo il 75% di precarico per 25 cicli (P < .01).

*Conclusioni:* Il precarico torsionale ha ridotto la resistenza alla fatica ciclica di strumenti in NiTi convenzionale e M-wire (come Mtwo e ProTaper Next) ad eccezione degli Mtwo che hanno subito il 25% o il 50% di precarico torsionale.

 $\odot$  2015 Società Italiana di Endodonzia. Production and hosting by Elsevier B.V. Tutti i diritti riservati.

### Introduction

Unexpected nickel-titanium (NiTi) rotary instrument separation during use is a concern<sup>1-4</sup> and has a potential effect on the outcome of treatment.<sup>5</sup>

Many variables might contribute to file separation, but the 2 main causes are cyclic fatigue and torsional fatigue. Each has been adequately defined in endodontic literature<sup>6,7</sup> and clinically, cyclic fatigue seems to be more prevalent in curved root canals, whereas torsional failure might happen even in a straight canal.<sup>8</sup>

Although both these failure modes probably occur simultaneously in a clinical situation,<sup>6</sup> studies have found cyclic fatigue to be the primary cause of file separation.<sup>9</sup>

Torsional fracture occurs when the torque resulting from the contact between the instrument and canal wall exceeds the torsional strength of the instrument or when the instrument tip is locked in a canal while the rest continues to rotate.<sup>10,11</sup>

Instrument fractured by fatigue rotates freely in a curvature, generating tension/compression cycles at the point of maximum flexure until the fracture occurs.<sup>12</sup>

Manufacturers are constantly attempting to improve files. M-wire was recently introduced as innovative systems that have a higher cyclic fatigue resistance than conventional NiTi. The M-wire NiTi, as the novel file system ProTaper Next (Dentsply Maillefer, Ballaigues, Switzerland), is subjected to thermomechanical processing resulting in a reported increased flexibility.  $^{13,\,14}$ 

Many fracture simulation studies of NiTi files have been conducted separately for a cyclic fatigue or torsional failure test.<sup>6,15</sup> Only few studies have tried to correlate these 2 factors of fracture.<sup>16–19</sup> Some reports<sup>16,17</sup> have found that the cyclic preloading (preuse) of NiTi rotary files would reduce the torsional resistance significantly. However, to date, few studies with contrasting findings have investigated the potential effect on cyclic fatigue resistance by torsional preloading.<sup>18,19</sup> In particular, no studies have investigated the influence of torsional preloading on cyclic fatigue resistance of endodontic rotary instruments made by traditional NiTi or M-Wire as Mtwo (Sweden & Martina, Due Carrare, PD, Italy) and ProTaper Next.

The aim of this study was to evaluate the effect of different torsional preload on the cyclic fatigue resistance of NiTi rotary instruments made by NiTi or M-Wire.

#### Materials and methods

A total of 160, 25-mm-long, new files Mtwo and ProTaper Next X2, eighty for each brand and all size #25 with a 0.06 taper were tested. Ten instruments for every brand were torsionally tested until fracture to estimate their mean ultimate torsional strength. Four torsional preloading conditions were



**Figure 1** Custom-made torsional device produced following the ISO 3630-1. (A) ProTaper Next X2 clumped at 5 mm from the tip by the chuck. (B) Mtwo #25/.06 instrument broken at 5 mm from the tip after the cyclic fatigue test using an artificial canal.

applied to the remaining instruments of each brand: 0% or no preloading (n = 10) and 25%, 50%, and 75% of the mean ultimate torsional strength (n = 20 each). The twenty files tested for every extent of torsional preloading were subdivided in two subgroups (n = 10) on the basis of the number of preloading cycles performed (25 or 50 torsional cycles). The torsional load until fracture and established torsional preloads were applied using the same custom-made device produced following the ISO 3630-1 (Fig. 1A).

Each file was clamped at 5 mm from the tip using a chuck connected to a torque-sensing load cell, after which the shaft of the file was fastened into an opposing chuck able to be rotated with a stepper motor. Each file shaft was rotated in the clockwise direction at a speed of 2 revolutions per minute until file separation or the preset torque value (25%, 50%, and 75%). When the preset torque value was achieved, the motor automatically moved in counterclockwise direction, at the same rotational rate, until the instrument returned in the original position (i.e., strain = 0). This motion was repeated until the preset number of repetitions (25 or 50) was reached. The torque load (Ncm) and angular rotation (°, degrees) were monitored continuously by using a torsiometer (Sabri Dental Enterprises, Downers Grove, IL) at room temperature (20 °C  $\pm$  1 °C), and the ultimate torsional strength and angle of rotation at failure were recorded.

After torsional preloading, a static model for cyclic fatigue testing was conducted in an artificial curved canal of a custom-made device already used in many published studies.<sup>8,12,15,20</sup> The artificial canal was manufactured by reproducing the instrument's size and taper and it had 5 mm radius of curvature (measured at the internal concave surface of the artificial canal), 60° angle of curvature measured according to the Schneider's method,<sup>21</sup> a center of curvature at 5 mm from the end of the canal and a curved segment of 5 mm in length (Fig. 1B). The files were activated by using a 6:1 reduction handpiece (Sirona Dental Systems GmbH, Bensheim, Germany) powered by a torque-controlled motor (Silver Reciproc; VDW, Munich, Germany) at a speed of 300 rpm. To reduce friction between the instrument and the metal canal walls, a special high-flow synthetic oil designed for the lubrication of mechanical parts (Super Oil; Singer Co Ltd, Elizabethport, NJ) was applied. All instruments were rotated until a fracture occurred. The time was recorded and stopped when a fracture was detected visually and/or audibly.

The number of cycles to failure (NCF) for each instrument was recorded. The length of the fractured file tip was measured by using a digital microcaliper (Mitutoyo Italiana srl, Lainate, Italy).

The fracture surfaces of all fragments were examined under a Scanning Electron Microscope (SEM) (ZEISS Supra 35VP, Oberkochen, GmBH, Germany) looking for topographic features of the fractured instruments.

NCF data were analyzed by using 2-way analysis of variance and the Bonferroni post hoc test (Prism 5.0; GraphPad Software, Inc., La Jolla, CA) at 0.05 as level of significance.

#### Results

The mean torque load and angle of rotation until fracture are presented in Table 1.

Every amount (75%, 50% and 25%) of torsional preloading significantly decreased the cyclic fatigue resistance of all nonpreloaded (0%) ProTaper Next X2 either after 50 or after 25 cycles (P < 0.0001 and < 0.001 respectively). Every amount of torsional preloading reduced cyclic fatigue resistance of the new Mtwo 25/0.06 after 50 cycles (P < 0.01). However, only Mtwo instruments preloaded at 75% of the maximum torsional strength showed a reduced cyclic fatigue after 25 cycles (P < 0.05) (Table 2).

 Table 1
 Mean torque (Ncm) and angle of rotation (°) of instruments tested.

| Instrument  | Torque (Ncm)      |                    |      |      | Angle of rotation ( $^{\circ}$ ) |                    |     |     |  |
|---|-------------------|--------------------|------|------|----------------------------------|--------------------|-----|-----|--|
|   | Mean              | Standard deviation | Min  | Max  | Mean                             | Standard deviation | Min | Max |  |
| ProTaper Next X2  | 1.30 <sup>ª</sup> | 0.20               | 0.92 | 1.70 | 262.24 <sup>c</sup>              | 20.61              | 220 | 285 |  |
| Mtwo #25/.06  | 0.94 <sup>b</sup> | 0.15               | 0.77 | 1.18 | 348.73 <sup>c</sup>              | 71.43              | 235 | 430 |  |
| Different superscript letters indicate statistic differences among groups ( $P < 0.05$ ). |                   |                    |      |      |                                  |                    |     |     |  |

| File             | Without torsional preloading (0%) | RNTP     | Torsional preloading  |  |  |
|------------------|-----------------------------------|----------|---|--|--|
|                  |                                   |          | 25%   | 50%  | 75%  |
| ProTaper Next X2 | $405\pm 30^{a}$                   | 25<br>50 | $\begin{array}{c} 290 \pm 20^{\text{c}} \\ 212 \pm 51^{\text{c}} \end{array}$ | $272 \pm 53^{c}$<br>$220 \pm 90^{c}$   | $\begin{array}{c} 265 \pm 37^{c} \\ 203 \pm 65^{c} \end{array}$  |
| Mtwo             | $606\pm26^{b}$                    | 25<br>50 | $\begin{array}{c} 510\pm47^{d}\\ 398\pm81^{d} \end{array}$                    | $\begin{array}{c} \text{502} \pm \text{122}^{\text{d}} \\ \text{403} \pm \text{93}^{\text{d}} \end{array}$ | $\begin{array}{c} \text{427} \pm \text{101}^{\text{d}} \\ \text{352} \pm \text{98}^{\text{d}} \end{array}$ |

Table 2 Number of cycles to failure (NCF) (mean  $\pm$  standard deviation) after different percentage and cycles of torsional preloading.

Different superscript letters indicate statistic differences among groups (P < 0.05).



**Figure 2** SEM appearances of the torsionally preloaded instruments (75% for 50 cycles) after fatigue test (A,a: Mtwo; B, b: ProTaper Next). (A, B) Fracture surface of preloaded files at 75% of the ultimate torsional strength for 50 cycles and their lateral view (a, b). Dimples involve all fracture surface of torsionally preloaded instruments fractured by cyclic fatigue test. Smoothed borders (white arrows) due to the torsional preload are visible in lateral view.

The length of the fractured files was not statistically different for all of the instruments tested (P > 0.05) and it ranged from 4.5 to 6.2 mm (mean = 5.2 mm).

SEM of the fracture surface showed similar features of torsional failure with fibrous dimples at the center of rotation for instruments tested only at torsion. Dimples are visible in all fracture surfaces of the instruments tested for cyclic fatigue. Dimples in the all fracture surface (Fig. 2A and B) and smoothed borders (Fig. 2a and b) were found in the torsional preloaded files broken in the following cyclic fatigue test.

#### Discussion

Endodontic instruments are subjected to both torsional and flexural stresses during root canal preparation and these types of stress can lead to metal fatigue and failure.<sup>18</sup>

There are only few studies that have investigated either the effect of cyclic fatigue preloading on the torsional resistance<sup>16,17,22</sup> or the effect of torsional preloads on the cyclic fatigue resistance of conventional NiTi instruments. 18,19,23 However, no data have been reported on the influence of torsional preloading on cyclic fatigue of instruments made by different NiTi alloys. Therefore, in this study, M-wire (ProTaper Next) and conventional NiTi files (Mtwo) were compared for the effect of different torsional preloads on their cyclic fatigue resistance.

NCF value of instruments, tested in custom made devices, is usually used to investigate cyclic fatigue resistance.<sup>24–26</sup> On the other hand, the ISO standard  $3630-1^{16,18,23}$  is often used to measure torsional strength of NiTi instruments. In this study, cyclic fatigue was measured by NCF testing the files in a custom made device already used in previous study,<sup>8,15,25</sup> while the torsional loads were applied following the ISO Standard 3630-1 modified only for clamping the testing instruments at the D5 than at the D3 level. This modification was performed in order to provide the torsional preloads at the same level of the maximum stress of the subsequent cyclic fatigue where it was used an artificial canal with 5 mm of radius of curvature.

The broken fragment after the fatigue tests showed an average length of 5 mm that coincided with the site of torsional preloading application (at D5).

Although the comparison among the different instruments is difficult to make because of their differences in design and cross-section, in this study, new files without torsional preloads Mtwo showed significantly higher cyclic fatigue resistance than ProTaper Next (P < 0.001). On the other hand, new ProTaper Next X2 showed significantly higher value of the mean ultimate torsional strength than Mtwo #25/0.06 (P < 0.001). These results are in agreement with a recent study that has reported higher cyclic fatigue resistance of Mtwo than ProTaper Next.<sup>27</sup> Even if no previous studies have compared the torque resistance of Mtwo and ProTaper Next, recently it was reported that instruments made by M-wire as ProTaper Next have higher torsional resistance than instruments made by conventional NiTi as Mtwo<sup>28</sup> (Table 1).

Analyzing the influence of torsion on cyclic fatigue, in agreement with other reports in literature, the number of torsional preload repetitions used (25 or 50 cycles) did not affect the cyclic fatigue resistance of the ProTaper Next after every amount and Mtwo after the 75% of torsional preload.<sup>19</sup> However, Mtwo torsionally preloaded up to the 50% of their maximum torsional strength reduce their cyclic fatigue when 50 cycles was performed, but no after 25 cycles. This finding is probably due to the cross-section and design of Mtwo instruments that are capable to not reduce their cyclic fatigue resistance after a medium torsional stress (up to 50%) applied for a reduced number of cycles (25 cycles).

On the other hand, the extent of torsional pre-loads reduced the cyclic fatigue resistance of the instruments tested especially when 75% of the mean ultimate torsional strength was applied. These results are in agreement with other studies in literature that have shown a reduced cyclic fatigue resistance of conventional NiTi instruments, as K3, after their torsional pre-load.<sup>18,23</sup> However, another recent study has shown that the torsional preloads may improve the cyclic fatigue resistance of conventional nickel-titanium rotary instruments as ProFile and ProTaper F1.<sup>19</sup> These contrasting findings are probably due to the different instruments and methodology used. Moreover in this last study torsional preloads were applied at 5 mm from the tip, but the cyclic fatigue test was performed in an artificial canal with 7.8 mm of radius of curvature; thus, the two stresses were performed in different parts of instruments making it difficult to draw conclusions on the influence of torsional stress on cyclic fatigue of the tested instruments.

The SEM analysis showed typical fractographic appearances of torsional and/or cyclic fatigue fractures similar among the 2 brands tested (Fig. 2). Dimples on the entire fracture surface characterize cyclic fatigue fracture, while torsional failure is characterized by circular abrasion marks and dimples near the center of rotation on the fracture surface.  $^{11,17,29}$ 

In this study, when the cyclic fatigue test was applied after the incomplete torsional test, instruments of all brands showed typical cyclic fatigue fracture pattern at the SEM evaluation. Therefore, in preloaded instruments, as previously reported,<sup>17</sup> the fractographic findings corresponded to those ones due to the last test performed (cyclic fatigue in this report).

#### Conclusions

Within the limitations of this study, it could be concluded that torsional preloads reduced the cyclic fatigue resistance of conventional and M-wire NiTi rotary instruments except for Mtwo #25/0.06 preloaded by a medium torsional stress for few cycles (25% or 50% of torsional preloading applied for 25 cycles).

#### **Clinical relevance**

ProTaper Next are able to resist to high torsional stress whenever they are used in narrow root canals, while Mtwo endures high flexural stress whenever they are used in curved root canals.

#### **Conflict of interest**

The authors deny any conflicts of interest.

#### Acknowledgements

The authors thank engineers Salvatore Caltabiano and Salvatore Franco from University of Catania for the support in the torsional tests.

#### References

- Cheung GS, Liu CSY. A retrospective study of endodontic treatment outcome between nickel-titanium rotary and stainless steel hand filing techniques. J Endod 2009;35:938–43.
- Walia HM, Brantley WA, Gerstein H. An initial investigation of the bending and torsional properties of nitinol root canal files. J Endod 1988;14:346-51.
- Shen Y, Riyahi AM, Campbell L, Zhou H, Du T, Whang Z, et al. Effect of a combination of torsional and cyclic fatigue preloading on the fracture behavior of K3 and K3XF instruments. *J Endod* 2015;41:526–30.
- Cheung GS. Instrument fracture: mechanisms, removal of fragments, and clinical outcomes. Endod Top 2009;16:1–26.
- Pedullà E, Grande NM, Plotino G, Palermo F, Gambarini G, Rapisarda E. Cyclic fatigue resistance of two reciprocating nickel-titanium instruments after immersion in sodium hypochlorite. Int Endod J 2013;46:155–9.
- 6. Yum J, Cheung GSP, Park JK, Hur B, Kim HC. Torsional strength and toughness of nickel-titanium rotary files. *J Endod* 2011;37:382-6.
- Bhagabati N, Yadav S, Talwar S. An in vitro cyclic fatigue analysis of different endodontic nickel-titanium rotary instruments. J Endod 2012;38:515-8.
- Plotino G, Grande NM, Melo MC, Bahia MG, Testarelli L, Gambarini G. Cyclic fatigue of NiTi rotary instruments in a simulated apical abrupt curvature. *Int Endod J* 2010;43:226–30.
- Parashos P, Gordon I, Messer HH. Factors influencing defects of rotary nickeltitanium files after clinical use. J Endod 2004;30:722-5.
- Peters OA, Barbakow F. Dynamic torque and apical forces of ProFile .04 rotary instruments during preparation of curved canals. *Int Endod J* 2002;35:379–89.
- Parashos P, Messer HH. Rotary NiTi instrument fracture and its consequences. J Endod 2006;32:1031–43.
- 12. Pedullà E, Grande NM, Plotino G, Gambarini G, Rapisarda E. Influence of continuous or reciprocating motion on cyclic fatigue

resistance of 4 different nickel-titanium rotary instruments. *J* Endod 2013;**39**:258-61.

- **13.** Alapati SB, Brantley WA, Iijima M, Clark WA, Kovarik L, Buie C, et al. Metallurgical characterization of a new nickel-titanium wire for rotary endodontic instruments. *J Endod* 2009;**35**: 1589–93.
- Ninan E, Berzins DW. Torsion and bending properties of shape memory and superelastic nickel-titanium rotary instruments. J Endod 2013;39:101-4.
- Pedullà E, Plotino G, Grande NM, Pappalardo A, Rapisarda E. Cyclic fatigue resistance of four nickel-titanium rotary instruments: a comparative study. Ann Stomatol 2012;3:59–63.
- Kim JY, Cheung GS, Park SH, Ko DC, Kim JW, Kim HC. Effect from cyclic fatigue of nickel-titanium rotary files on torsional resistance. J Endod 2012;38:527–30.
- Campbell L, Shen Y, Zhou HM, Haapasalo M. Effect of fatigue on torsional failure of nickel-titanium controlled memory instruments. J Endod 2014;40:562–5.
- Bahia MG, Melo MC, Buono VT. Influence of cyclic torsional loading on the fatigue resistance of K3 instruments. *Int Endod* J 2008;10:883–91.
- Cheung GS, Oh SH, Ha JH, Kim SK, Park SH, Kim HC. Effect of torsional loading of nickel-titanium instruments on cyclic fatigue resistance. J Endod 2013;39:1593–7.
- Al-Sudani D, Grande NM, Plotino G, Pompa G, Di Carlo S, Testarelli L, et al. Cyclic fatigue of nickel-titanium rotary instruments in a double (S-shaped) simulated curvature. J Endod 2012;38:987–9.

- 21. Schneider SW. A comparison of canal preparations in straight and curved root canals. *Oral Surg* 1971;32:271–5.
- 22. Ullmann CJ, Peters OA. Effect of cyclic fatigue on static fracture loads in ProTaper nickel—titanium rotary instruments. *J Endod* 2005;**31**:183-6.
- Galvão Barbosa FO, Ponciano Gomes JA, Pimenta de Araújo MC. Influence of previous angular deformation on flexural fatigue resistance of K3 nickel-titanium rotary instruments. J Endod 2007;33:1477–80.
- Yao JH, Schwartz SA, Beeson TJ. Cyclic fatigue of three types of rotary nickel-titanium files in a dynamic model. J Endod 2006;32:55–7.
- Pedullà E, Plotino G, Grande NM, Scibilia M, Pappalardo A, Malagnino VA, et al. Influence of rotational speed on the cyclic fatigue of Mtwo instruments. *Int Endod J* 2014;47:514–9.
- Lopes HP, Ferreira AA, Elias CN, Moreira EJ, de Oliveira JC, Siqueira Jr JF. Influence of rotational speed on the cyclic fatigue of rotary nickel—titanium endodontic instruments. *J Endod* 2009;35:1013–6.
- 27. Chaveli-Diaz B, Forner L, Llena C, Madureira R, Tadeu F. Cyclic fatigue resistance of five endodontic instruments. In: Natal Jorge, et al., editors. *Biodental engineering III*. London: Taylor and Francis Group; 2014. ISBN 978-1-138-02671-1.
- Elnaghy AM, Elsaka SE. Torsion and bending properties of One-Shape and WaveOne instruments. J Endod 2015;41:544–7.
- 29. Pedullà E, Franciosi G, Ounsi HF, Tricarico M, Rapisarda E, Grandini S. Cyclic fatigue resistance of nickel-titanium instruments after immersion in irrigant solutions with or without surfactants. *J Endod* 2014;40:1245–9.